



NOTES:

1. Connection to Data Acquisition System (DAS) to be made using conduit or watertight flex into ELECTRICAL INTERFACE ENCLOSURE, following diagram detail.
2. Single 15 amp minimum service provides power to all components.
3. Suggested cord/flex routing. Actual will be site specific.

ChemScan[®]
UV-2150 Analyzer System Installation Diagram
- Front and Rear Views -
Stand Mount

ASA, Inc.	PHONE: 414-650-2280 FAX: 414-650-2285
Applied Spectrometry Associates, Inc. © 98	
DATE: 6-11-98 BY: LJB	
ChemScan UV-2150 Analyzer System Installation Diagram - Stand	
DRAWING #: A980611-143	
FILE # 21STNDF1	